

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
 Data File : VN084808.D
 Acq On : 12 Nov 2024 20:02
 Operator : JC\MD
 Sample : P4809-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-5

Quant Time: Nov 13 02:37:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	174663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	125511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	120128	47.618	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.240%
35) Dibromofluoromethane	8.165	113	97082	46.867	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	10.565	98	365253	47.867	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.740%
62) 4-Bromofluorobenzene	12.847	95	143146	50.195	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.400%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	35323	46.147	ug/l	90
20) Methylene Chloride	5.259	84	5424	2.445	ug/l #	61
43) Isopropyl Acetate	8.688	43	107480	21.776	ug/l #	88
95) Naphthalene	15.635	128	15965	2.198	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111224\
Data File : VN084808.D
Acq On : 12 Nov 2024 20:02
Operator : JC\MD
Sample : P4809-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-5

Quant Time: Nov 13 02:37:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

