

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111225\
 Data File : VN088266.D
 Acq On : 12 Nov 2025 10:41
 Operator : JC\MD
 Sample : VN1112MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1112MBL01

Quant Time: Nov 14 00:09:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	195618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	433845	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	420389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	198881	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	193619	57.239	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.480%
35) Dibromofluoromethane	8.147	113	147454	50.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	10.547	98	511876	45.579	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.160%
62) 4-Bromofluorobenzene	12.829	95	187331	47.232	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111225\
Data File : VN088266.D
Acq On : 12 Nov 2025 10:41
Operator : JC\MD
Sample : VN1112MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1112MBL01

Quant Time: Nov 14 00:09:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

