

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111318\
 Data File : VN052288.D
 Acq On : 13 Nov 2018 15:57
 Operator : MD\SY
 Sample : PB114744ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744ZHE#01

Quant Time: Nov 14 01:25:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	340530	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	564140	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	464480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147728	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	220827	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
35) Dibromofluoromethane	7.59	113	179321	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	10.09	98	690390	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
62) 4-Bromofluorobenzene	12.40	95	190777	41.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.50%	
Target Compounds						
20) Methylene Chloride	4.55	84	16796	3.956	ug/l	98
43) Isopropyl Acetate	8.17	43	128481	17.211	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111318\
Data File : VN052288.D
Acq On : 13 Nov 2018 15:57
Operator : MD\SY
Sample : PB114744ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114744ZHE#01

Quant Time: Nov 14 01:25:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

