

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111318\
 Data File : VN052292.D
 Acq On : 13 Nov 2018 17:43
 Operator : MD\SY
 Sample : PB114744ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744ZHE#05

Quant Time: Nov 14 02:27:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	325588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	546260	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	466219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	158269	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	220700	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
35) Dibromofluoromethane	7.59	113	173399	49.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.02%	
50) Toluene-d8	10.09	98	682604	49.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.58%	
62) 4-Bromofluorobenzene	12.40	95	196511	43.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.76%	
Target Compounds						
20) Methylene Chloride	4.56	84	17078	4.207	ug/l	95
43) Isopropyl Acetate	8.17	43	134340	18.585	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111318\
Data File : VN052292.D
Acq On : 13 Nov 2018 17:43
Operator : MD\SY
Sample : PB114744ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114744ZHE#05

Quant Time: Nov 14 02:27:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

