

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111318\
 Data File : VN052296.D
 Acq On : 13 Nov 2018 19:30
 Operator : MD\SY
 Sample : PB114744ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744ZHE#09

Quant Time: Nov 14 02:39:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	307214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	521347	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	436780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153565	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	211916	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
35) Dibromofluoromethane	7.59	113	164773	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
50) Toluene-d8	10.09	98	644337	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	12.40	95	184791	43.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.48%	
Target Compounds						
20) Methylene Chloride	4.55	84	10166	2.654	ug/l	92
43) Isopropyl Acetate	8.17	43	171257	24.824	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111318\
Data File : VN052296.D
Acq On : 13 Nov 2018 19:30
Operator : MD\SY
Sample : PB114744ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114744ZHE#09

Quant Time: Nov 14 02:39:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

