

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111318\
 Data File : VN052300.D
 Acq On : 13 Nov 2018 21:17
 Operator : MD\SY
 Sample : PB114744ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744ZHE#13

Quant Time: Nov 14 02:51:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	322789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	541006	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	448925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	219881	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
35) Dibromofluoromethane	7.59	113	166901	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	10.09	98	673265	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	12.40	95	188731	42.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.10%	
Target Compounds						
20) Methylene Chloride	4.56	84	7771	1.931	ug/l	95
43) Isopropyl Acetate	8.17	43	176201	24.613	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111318\
Data File : VN052300.D
Acq On : 13 Nov 2018 21:17
Operator : MD\SY
Sample : PB114744ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114744ZHE#13

Quant Time: Nov 14 02:51:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

