

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111318\
 Data File : VN052286.D
 Acq On : 13 Nov 2018 15:03
 Operator : MD\SY
 Sample : PB114744TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744TB

Quant Time: Nov 14 01:20:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	329195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	553431	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	462173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	157200	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	219482	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
35) Dibromofluoromethane	7.59	113	178021	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
50) Toluene-d8	10.09	98	684735	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	12.40	95	190280	41.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.88%	
Target Compounds						
20) Methylene Chloride	4.56	84	8549	2.083	ug/l	89
43) Isopropyl Acetate	8.17	43	169831	23.190	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111318\
Data File : VN052286.D
Acq On : 13 Nov 2018 15:03
Operator : MD\SY
Sample : PB114744TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114744TB

Quant Time: Nov 14 01:20:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052286.D

Acq: 13 Nov 2018 15:03

Instrument :

MSVOA_N

ClientSampleId :


PB114744TB

Tot Ion:168 Resp: 329195

Ion Ratio Lower Upper

168 100

99 61.1 44.7 67.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

150000



100000

50000

0

Time-->

7.60 7.70 7.80



#20

Methylene Chloride

Concen: 2.083 ug/l

RT: 4.56 min Scan# 866

Delta R.T. 0.00 min

Lab File: VN052286.D

Acq: 13 Nov 2018 15:03

Tgt Ion: 84 Resp: 8549


Ion Ratio Lower Upper

84 100

49 126.4 114.8 172.2

51 44.5 33.1 49.7

86 57.1 49.7 74.5



Abundance Ion 83.90 (83.60 to 84.60): VNO

Ion 48.90 (48.60 to 49.60): VNO

Ion 50.90 (50.60 to 51.60): VNO

Ion 85.90 (85.60 to 86.60): VNO

4.56

6000

5000

4000

3000


2000


1000

0

Time-->

4.45 4.50 4.55 4.60 4.65









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN052286.D
 Acq: 13 Nov 2018 15:03




Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744TB


Tot Ion: 65 Resp: 219482
 Ion Ratio Lower Upper
 65 100
 67 53.7 0.0 106.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN052286.D
 Acq: 13 Nov 2018 15:03

Tgt Ion: 114 Resp: 553431
 Ion Ratio Lower Upper
 114 100
 63 22.4 0.0 43.4
 88 15.6 0.0 30.2









#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN052286.D
 Acq: 13 Nov 2018 15:03

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114744TB

Tot Ion:113 Resp: 178021
 Ion Ratio Lower Upper
 113 100
 111 101.6 80.2 120.4
 192 17.2 14.1 21.1




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#43
 Isopropyl Acetate
 Concen: 23.190 ug/l
 RT: 8.17 min Scan# 1988
 Delta R.T. 0.00 min
 Lab File: VN052286.D
 Acq: 13 Nov 2018 15:03

Tgt Ion: 43 Resp: 169831
 Ion Ratio Lower Upper
 43 100
 61 17.2 14.2 21.4
 87 10.3 8.4 12.6



Abundance Ion 43.00 (42.70 to 43.70): VNO
 Ion 61.00 (60.70 to 61.70): VNO
 Ion 87.00 (86.70 to 87.70): VNO





#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052286.D

Acq: 13 Nov 2018 15:03

Instrument :

MSVOA_N

ClientSampleId :


PB114744TB

Tot Ion: 98 Resp: 684735


Ion Ratio Lower Upper

98 100



100 61.9 50.2 75.2



Abundance Ion 98.00 (97.70 to 98.70): VN052227.D (-2570) (-)



Time-->



#62

4-Bromofluorobenzene

Concen: Below ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN052286.D

Acq: 13 Nov 2018 15:03


Tgt Ion: 95 Resp: 190280

Ion Ratio Lower Upper


95 100

174 79.3 0.0 159.4


176 76.2 0.0 157.8




Abundance Ion 94.90 (94.60 to 95.60): VN052227.D (-3291) (-)



Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN052286.D
Acq: 13 Nov 2018 15:03

Instrument :
MSVOA_N
ClientSampleId :
PB114744TB

Tot Ion:117 Resp: 462173
Ion Ratio Lower Upper
117 100
82 57.6 46.2 69.2
119 32.4 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN052286.D
Acq: 13 Nov 2018 15:03

Tgt Ion:152 Resp: 157200
Ion Ratio Lower Upper
152 100
115 59.9 29.8 89.5
150 157.4 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

