

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111319\
 Data File : VN059210.D
 Acq On : 13 Nov 2019 11:09
 Operator : JC/SP
 Sample : K5799-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LTM-RB-6-SRI4

Quant Time: Nov 14 08:09:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	457984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	709320	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	602943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	205580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	296646	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
35) Dibromofluoromethane	7.56	113	226722	53.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.06%	
50) Toluene-d8	10.08	98	874318	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
62) 4-Bromofluorobenzene	12.40	95	270277	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
Target Compounds						
16) Acetone	3.79	43	647902	221.308	ug/l	100
25) 2-Butanone	6.81	43	18221	4.722	ug/l	95
29) Tetrahydrofuran	7.18	42	11544	4.709	ug/l	99
37) Ethyl Acetate	6.91	43	7857	1.065	ug/l #	81
67) Ethyl Benzene	11.51	91	21434	1.028	ug/l	97
68) m/p-Xylenes	11.61	106	25248	3.232	ug/l	95
69) o-Xylene	11.94	106	8940	1.184	ug/l	96
70) Styrene	11.96	104	10395	0.876	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111319\
Data File : VN059210.D
Acq On : 13 Nov 2019 11:09
Operator : JC/SP
Sample : K5799-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LTM-RB-6-SRI4

Quant Time: Nov 14 08:09:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

