


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111319\
Data File : VN059204.D
Acq On : 13 Nov 2019 8:11
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Title : SW846 8260
Last Update : Thu Nov 07 04:00:41 2019



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	15129	PASS
75	95	30	60	49.9	40034	PASS
95	95	100	100	100.0	80152	PASS
96	95	5	9	6.9	5563	PASS
173	174	0.00	2	0.6	389	PASS
174	95	50	100	75.0	60152	PASS
175	174	5	9	7.2	4316	PASS
176	174	95	101	98.1	58992	PASS
177	176	5	9	7.1	4204	PASS