

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111320\
 Data File : VN064631.D
 Acq On : 13 Nov 2020 15:00
 Operator : JC/MD
 Sample : L4747-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB178-GW-598-600

Manual Integrations
 APPROVED

MMDadoda
 11/17/2020 6:41:32 PM

Quant Time: Nov 17 02:47:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	219452	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	462985	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	477576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	169175	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	228131	59.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.36%	
35) Dibromofluoromethane	7.55	113	165124	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
50) Toluene-d8	10.06	98	596921	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
62) 4-Bromofluorobenzene	12.38	95	217835	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
Target Compounds						
16) Acetone	3.79	43	8920	6.126	ug/l	99
18) Methyl Acetate	4.28	43	2832m	0.989	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111320\
Data File : VN064631.D
Acq On : 13 Nov 2020 15:00
Operator : JC/MD
Sample : L4747-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB178-GW-598-600

Manual Integrations
APPROVED

MMDadoda
11/17/2020 6:41:32 PM

Quant Time: Nov 17 02:47:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

