

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111320\
 Data File : VN064644.D
 Acq On : 13 Nov 2020 20:46
 Operator : JC/MD
 Sample : L4719-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-D

Manual Integrations
 APPROVED

MMDadoda
 11/17/2020 6:41:49 PM

Quant Time: Nov 17 03:25:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	216398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	465061	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	516329	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	221674	58.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.62%	
35) Dibromofluoromethane	7.55	113	161365	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
50) Toluene-d8	10.06	98	618815	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.38	95	232047	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
Target Compounds						
16) Acetone	3.78	43	13206	9.198	ug/l	99
18) Methyl Acetate	4.28	43	4651m	1.648	ug/l	
20) Methylene Chloride	4.51	84	12508	4.391	ug/l	98
43) Isopropyl Acetate	8.13	43	83846	10.531	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111320\
Data File : VN064644.D
Acq On : 13 Nov 2020 20:46
Operator : JC/MD
Sample : L4719-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-D

Manual Integrations
APPROVED

MMDadoda
11/17/2020 6:41:49 PM

Quant Time: Nov 17 03:25:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

