

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111320\
 Data File : VN064646.D
 Acq On : 13 Nov 2020 21:35
 Operator : JC/MD
 Sample : L4705-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 32785

Manual Integrations
 APPROVED

MMDadoda
 11/17/2020 6:41:51 PM

Quant Time: Nov 17 03:29:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	216638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	479072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	524865	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183975	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	227309	60.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.48%	
35) Dibromofluoromethane	7.55	113	166035	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
50) Toluene-d8	10.06	98	627200	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.38	95	235978	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
Target Compounds						
16) Acetone	3.78	43	25467m	17.717	ug/l	
18) Methyl Acetate	4.29	43	4712m	1.667	ug/l	
20) Methylene Chloride	4.50	84	11960	4.194	ug/l #	77
43) Isopropyl Acetate	8.13	43	86661	10.566	ug/l	99
52) Toluene	10.12	92	112536	12.407	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111320\
Data File : VN064646.D
Acq On : 13 Nov 2020 21:35
Operator : JC/MD
Sample : L4705-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
32785

Manual Integrations
APPROVED

MMDadoda
11/17/2020 6:41:51 PM

Quant Time: Nov 17 03:29:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

