

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
 Data File : VN084818.D
 Acq On : 13 Nov 2024 11:13
 Operator : JC\MD
 Sample : VN1113MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1113MBL02

Quant Time: Nov 14 00:31:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	181889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	317168	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124016	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128884	49.060	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.120%
35) Dibromofluoromethane	8.165	113	105479	49.132	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	10.565	98	359892	45.508	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.020%
62) 4-Bromofluorobenzene	12.847	95	133391	45.132	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
Data File : VN084818.D
Acq On : 13 Nov 2024 11:13
Operator : JC\MD
Sample : VN1113MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1113MBL02

Quant Time: Nov 14 00:31:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

