

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
 Data File : VN084820.D
 Acq On : 13 Nov 2024 12:01
 Operator : JC\MD
 Sample : P4793-01
 Misc : 1.35g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M00-24-00345

Quant Time: Nov 14 00:32:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	165663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	283775	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	111342	46.533	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.060%		
35) Dibromofluoromethane	8.165	113	90017	46.864	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.720%		
50) Toluene-d8	10.565	98	321435	45.428	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.860%		
62) 4-Bromofluorobenzene	12.847	95	105229	39.793	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	79.580%		
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.588	83	8692	3.207	ug/l	93
52) Toluene	10.624	92	897967	179.469	ug/l	99
67) Ethyl Benzene	11.965	91	4880	0.549	ug/l #	57
68) m/p-Xylenes	12.071	106	5002	1.488	ug/l	95
69) o-Xylene	12.400	106	2063	0.656	ug/l	85
70) Styrene	12.412	104	1763	0.321	ug/l	94
73) Isopropylbenzene	12.694	105	1956	0.265	ug/l	86
77) Bromobenzene	12.982	156	799	0.403	ug/l	85
78) n-propylbenzene	13.029	91	4843	0.561	ug/l	94
79) 2-Chlorotoluene	13.123	91	2351	0.417	ug/l	84
80) 1,3,5-Trimethylbenzene	13.176	105	3276	0.536	ug/l	90
82) 4-Chlorotoluene	13.223	91	3091	0.521	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	7172	1.138	ug/l	99
85) sec-Butylbenzene	13.612	105	2880	0.403	ug/l	97
87) 1,3-Dichlorobenzene	13.729	146	2245	0.581	ug/l	91
89) n-Butylbenzene	14.053	91	5022	0.917	ug/l #	86
93) 1,2,4-Trichlorobenzene	15.388	180	2418	1.189	ug/l #	78
95) Naphthalene	15.635	128	10285	1.690	ug/l	97
96) 1,2,3-Trichlorobenzene	15.835	180	2843	1.440	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
Data File : VN084820.D
Acq On : 13 Nov 2024 12:01
Operator : JC\MD
Sample : P4793-01
Misc : 1.35g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M00-24-00345

Quant Time: Nov 14 00:32:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

