

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
 Data File : VN084825.D
 Acq On : 13 Nov 2024 14:48
 Operator : JC\MD
 Sample : P4722-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2(0-6)

Quant Time: Nov 14 00:36:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	172472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	295434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	119240	47.867	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.740%
35) Dibromofluoromethane	8.159	113	96824	48.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	10.565	98	337965	45.879	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.760%
62) 4-Bromofluorobenzene	12.847	95	127349	46.257	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.520%
Target Compounds						
16) Acetone	4.436	43	6357	6.051	ug/l	98
20) Methylene Chloride	5.265	84	33928	15.486	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
Data File : VN084825.D
Acq On : 13 Nov 2024 14:48
Operator : JC\MD
Sample : P4722-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2(0-6)

Quant Time: Nov 14 00:36:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

