

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
 Data File : VN084831.D
 Acq On : 13 Nov 2024 17:11
 Operator : JC\MD
 Sample : P4780-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Nov 14 00:39:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	166310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115326	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118467	49.319	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	98.640%	
35) Dibromofluoromethane	8.165	113	95395	48.192	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.380%	
50) Toluene-d8	10.565	98	340825	46.741	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	93.480%	
62) 4-Bromofluorobenzene	12.847	95	124990	45.865	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.720%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
Data File : VN084831.D
Acq On : 13 Nov 2024 17:11
Operator : JC\MD
Sample : P4780-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Nov 14 00:39:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

