

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
 Data File : VN084838.D
 Acq On : 13 Nov 2024 19:59
 Operator : JC\MD
 Sample : P4732-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PPE-Grab

Quant Time: Nov 14 00:42:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	176971	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	305388	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	143495	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	123076	48.151	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.300%		
35) Dibromofluoromethane	8.165	113	97132	46.989	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.980%		
50) Toluene-d8	10.565	98	368952	48.453	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.900%		
62) 4-Bromofluorobenzene	12.847	95	148858	52.308	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	104.620%		
Target Compounds						
16) Acetone	4.442	43	20094	24.643	ug/l	100
20) Methylene Chloride	5.265	84	3301	1.468	ug/l	84
25) 2-Butanone	7.483	43	11570	9.702	ug/l	99
43) Isopropyl Acetate	8.688	43	107046	21.731	ug/l #	91
95) Naphthalene	15.635	128	178434	21.484	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111324\
Data File : VN084838.D
Acq On : 13 Nov 2024 19:59
Operator : JC\MD
Sample : P4732-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PPE-Grab

Quant Time: Nov 14 00:42:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

