

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111418\
 Data File : VN052318.D
 Acq On : 14 Nov 2018 17:35
 Operator : MD\SY
 Sample : J5164-09 2.5ppb
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 MDL-WATER-03-QT4-2018

Quant Time: Nov 15 04:34:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	246807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	435975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	374788	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	146966	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	186573	53.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.04%	
35) Dibromofluoromethane	7.59	113	142953	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
50) Toluene-d8	10.09	98	541094	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.40	95	164795	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
Target Compounds						
10) Methyl Iodide	3.96	142	8385	2.275	ug/l #	91
11) Tert butyl alcohol	4.79	59	2048	11.125	ug/l #	72
13) Acrolein	3.61	56	3819	34.327	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111418\
Data File : VN052318.D
Acq On : 14 Nov 2018 17:35
Operator : MD\SY
Sample : J5164-09 2.5ppb
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
MDL-WATER-03-QT4-2018

Quant Time: Nov 15 04:34:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

