

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111418\
 Data File : VN052320.D
 Acq On : 14 Nov 2018 18:28
 Operator : MD\SY
 Sample : J5164-08 10ppb
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 LOQ-WATER-02-QT4-2018

Quant Time: Nov 15 04:12:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	259091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	445521	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	393300	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	165410	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	186003	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
35) Dibromofluoromethane	7.59	113	144150	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
50) Toluene-d8	10.09	98	558006	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
62) 4-Bromofluorobenzene	12.40	95	176326	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
Target Compounds						
10) Methyl Iodide	3.96	142	36113	9.332	ug/l	100
11) Tert butyl alcohol	4.78	59	10367	53.645	ug/l	99
13) Acrolein	3.61	56	15874	71.487	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111418\
Data File : VN052320.D
Acq On : 14 Nov 2018 18:28
Operator : MD\SY
Sample : J5164-08 10ppb
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
LOQ-WATER-02-QT4-2018

Quant Time: Nov 15 04:12:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

