

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111418\
 Data File : VN052324.D
 Acq On : 14 Nov 2018 20:15
 Operator : MD\SY
 Sample : PB114787ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114787ZHE#06

Quant Time: Nov 15 03:47:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	318049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	529355	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	451254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147744	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	219003	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
35) Dibromofluoromethane	7.59	113	167239	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	10.09	98	665597	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.40	95	186485	42.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.40%	
Target Compounds						
16) Acetone	3.83	43	5991	4.844	ug/l	96
20) Methylene Chloride	4.55	84	12975	2.787	ug/l	91
43) Isopropyl Acetate	8.17	43	32547	4.380	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	9913	1.049	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111418\
Data File : VN052324.D
Acq On : 14 Nov 2018 20:15
Operator : MD\SY
Sample : PB114787ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114787ZHE#06

Quant Time: Nov 15 03:47:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

