

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111418\
 Data File : VN052325.D
 Acq On : 14 Nov 2018 20:42
 Operator : MD\SY
 Sample : PB114787ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114787ZHE#07

Quant Time: Nov 15 03:49:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	299016	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	505869	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	427661	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153500	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	210461	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
35) Dibromofluoromethane	7.59	113	160961	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	10.09	98	629096	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	12.40	95	187673	44.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.90%	
Target Compounds						
16) Acetone	3.82	43	5916	5.088	ug/l #	86
18) Methyl Acetate	4.33	43	3447	1.480	ug/l #	71
20) Methylene Chloride	4.55	84	8168	1.631	ug/l	87
43) Isopropyl Acetate	8.17	43	166863	23.500	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111418\
Data File : VN052325.D
Acq On : 14 Nov 2018 20:42
Operator : MD\SY
Sample : PB114787ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114787ZHE#07

Quant Time: Nov 15 03:49:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

