

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111418\
 Data File : VN052327.D
 Acq On : 14 Nov 2018 21:35
 Operator : MD\SY
 Sample : J5925-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5S

Quant Time: Nov 15 03:53:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	333265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	559721	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	483593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	183358	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	229277	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
35) Dibromofluoromethane	7.59	113	176618	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	10.09	98	705447	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	12.40	95	214301	45.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.74%	
Target Compounds						
16) Acetone	3.82	43	10250	7.909	ug/l	94
20) Methylene Chloride	4.55	84	28840	6.708	ug/l	97
43) Isopropyl Acetate	8.17	43	22825	2.905	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111418\
Data File : VN052327.D
Acq On : 14 Nov 2018 21:35
Operator : MD\SY
Sample : J5925-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-5S

Quant Time: Nov 15 03:53:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

