

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111419\
 Data File : VN059234.D
 Acq On : 14 Nov 2019 15:22
 Operator : JC/SP
 Sample : K5842-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S36-TB-2019-3

Manual Integrations
 APPROVED

MMDadoda
 11/15/2019 11:39:35 AM

Quant Time: Nov 15 07:48:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	466696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	720722	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	621442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	226068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	290703	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
35) Dibromofluoromethane	7.57	113	225624	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
50) Toluene-d8	10.08	98	871550	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.40	95	276054	45.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.34%	
Target Compounds						
30) Chloroform	7.36	83	3799m	0.420	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111419\
Data File : VN059234.D
Acq On : 14 Nov 2019 15:22
Operator : JC/SP
Sample : K5842-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S36-TB-2019-3

Manual Integrations
APPROVED

MMDadoda
11/15/2019 11:39:35 AM

Quant Time: Nov 15 07:48:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

