

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111419\
 Data File : VN059246.D
 Acq On : 14 Nov 2019 19:46
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 15 08:22:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.504	0.434	13.9	71	0.00
3 P	Chloromethane	0.685	0.595	13.1	74	0.00
4 C	Vinyl Chloride	0.664	0.603	9.2#	75	0.00
5 T	Bromomethane	0.382	0.314	17.8	69	0.00
6 T	Chloroethane	0.393	0.371	5.6	79	0.00
7 T	Trichlorofluoromethane	0.794	0.772	2.8	79	0.00
8 T	Diethyl Ether	0.323	0.333	-3.1	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.479	0.463	3.3	79	0.00
10 T	Methyl Iodide	0.593	0.508	14.3	68	0.00
11 T	Tert butyl alcohol	0.117	0.125	-6.8	87	0.00
12 CM	1,1-Dichloroethene	0.466	0.458	1.7#	80	0.00
13 T	Acrolein	0.060	0.070	-16.7	95	0.00
14 T	Allyl chloride	0.958	0.925	3.4	80	0.00
15 T	Acrylonitrile	0.292	0.326	-11.6	87	0.00
16 T	Acetone	0.320	0.266	16.9	62	0.00
17 T	Carbon Disulfide	1.448	1.202	17.0	70	0.00
18 T	Methyl Acetate	0.759	0.851	-12.1	90	0.00
19 T	Methyl tert-butyl Ether	1.606	1.706	-6.2	84	0.00
20 T	Methylene Chloride	0.557	0.554	0.5	82	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.494	1.4	80	0.00
22 T	Diisopropyl ether	1.810	1.888	-4.3	83	0.00
23 T	Vinyl Acetate	1.403	1.569	-11.8	83	0.00
24 P	1,1-Dichloroethane	1.011	1.024	-1.3	82	0.00
25 T	2-Butanone	0.421	0.434	-3.1	79	0.00
26 T	2,2-Dichloropropane	0.884	0.837	5.3	79	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.595	-2.4	83	0.00
28 T	Bromochloromethane	0.304	0.283	6.9	77	0.00
29 T	Tetrahydrofuran	0.268	0.292	-9.0	85	0.00
30 C	Chloroform	0.970	0.998	-2.9#	84	0.00
31 T	Cyclohexane	0.957	0.851	11.1	74	0.00
32 T	1,1,1-Trichloroethane	0.815	0.839	-2.9	82	0.00
33 S	1,2-Dichloroethane-d4	0.615	0.590	4.1	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.299	0.308	-3.0	85	0.00
36 T	1,1-Dichloropropene	0.467	0.466	0.2	80	0.00
37 T	Ethyl Acetate	0.520	0.557	-7.1	84	0.00
38 T	Carbon Tetrachloride	0.465	0.469	-0.9	82	0.00
39 T	Methylcyclohexane	0.542	0.519	4.2	75	0.00
40 TM	Benzene	1.407	1.435	-2.0	82	0.00
41 T	Methacrylonitrile	0.226	0.210	7.1	68	0.00
42 TM	1,2-Dichloroethane	0.483	0.489	-1.2	82	0.00
43 T	Isopropyl Acetate	0.829	0.876	-5.7	85	0.00
44 TM	Trichloroethene	0.346	0.349	-0.9	81	0.00
45 C	1,2-Dichloropropane	0.382	0.398	-4.2#	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111419\
 Data File : VN059246.D
 Acq On : 14 Nov 2019 19:46
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 15 08:22:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.238	0.242	-1.7	84	0.00
47 T	Bromodichloromethane	0.476	0.506	-6.3	84	0.00
48 T	Methyl methacrylate	0.371	0.390	-5.1	82	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	87	0.00
50 S	Toluene-d8	1.176	1.166	0.9	81	0.00
51 T	4-Methyl-2-Pentanone	0.469	0.565	-20.5	87	0.00
52 CM	Toluene	0.844	0.876	-3.8#	83	0.00
53 T	t-1,3-Dichloropropene	0.550	0.568	-3.3	82	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.614	-3.9	82	0.00
55 T	1,1,2-Trichloroethane	0.341	0.362	-6.2	86	0.00
56 T	Ethyl methacrylate	0.507	0.580	-14.4	86	0.00
57 T	1,3-Dichloropropane	0.588	0.621	-5.6	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.190	0.227	-19.5	84	0.00
59 T	2-Hexanone	0.360	0.416	-15.6	83	0.00
60 T	Dibromochloromethane	0.361	0.390	-8.0	86	0.00
61 T	1,2-Dibromoethane	0.346	0.368	-6.4	85	0.00
62 S	4-Bromofluorobenzene	0.424	0.439	-3.5	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.340	0.329	3.2	80	0.00
65 PM	Chlorobenzene	0.975	1.013	-3.9	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.390	-6.6	86	0.00
67 C	Ethyl Benzene	1.729	1.841	-6.5#	83	0.00
68 T	m/p-Xylenes	0.648	0.688	-6.2	83	0.00
69 T	o-Xylene	0.626	0.657	-5.0	84	0.00
70 T	Styrene	0.985	1.093	-11.0	83	0.00
71 P	Bromoform	0.284	0.312	-9.9	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.637	3.727	-2.5	84	0.00
74 T	N-amyl acetate	1.541	1.644	-6.7	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.239	1.286	-3.8	88	0.00
76 T	1,2,3-Trichloropropane	1.138	1.169	-2.7	92	0.00
77 T	Bromobenzene	0.914	0.896	2.0	83	0.00
78 T	n-propylbenzene	4.164	4.329	-4.0	83	0.00
79 T	2-Chlorotoluene	2.583	2.589	-0.2	85	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.150	-2.2	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.451	-4.6	86	0.00
82 T	4-Chlorotoluene	2.591	2.620	-1.1	83	0.00
83 T	tert-Butylbenzene	2.661	2.704	-1.6	84	0.00
84 T	1,2,4-Trimethylbenzene	2.982	3.173	-6.4	85	0.00
85 T	sec-Butylbenzene	3.461	3.623	-4.7	84	0.00
86 T	p-Isopropyltoluene	3.056	3.255	-6.5	83	0.00
87 T	1,3-Dichlorobenzene	1.565	1.590	-1.6	84	0.00
88 T	1,4-Dichlorobenzene	1.566	1.569	-0.2	84	0.00
89 T	n-Butylbenzene	2.675	2.761	-3.2	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111419\
Data File : VN059246.D
Acq On : 14 Nov 2019 19:46
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 15 08:22:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.614	0.607	1.1	86	0.00
91 T	1,2-Dichlorobenzene	1.539	1.588	-3.2	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.259	7.5	87	0.00
93 T	1,2,4-Trichlorobenzene	0.825	0.886	-7.4	85	0.00
94 T	Hexachlorobutadiene	0.531	0.531	0.0	86	0.00
95 T	Naphthalene	2.265	2.450	-8.2	86	0.00
96 T	1,2,3-Trichlorobenzene	0.863	0.890	-3.1	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6