


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111419\
Data File : VN059217.D
Acq On : 14 Nov 2019 8:13
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Title : SW846 8260
Last Update : Thu Nov 07 04:00:41 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	16285	PASS
75	95	30	60	52.0	44187	PASS
95	95	100	100	100.0	84898	PASS
96	95	5	9	6.7	5648	PASS
173	174	0.00	2	1.2	814	PASS
174	95	50	100	78.9	67016	PASS
175	174	5	9	7.3	4880	PASS
176	174	95	101	97.3	65178	PASS
177	176	5	9	6.4	4157	PASS