

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
 Data File : VN075266.D
 Acq On : 14 Nov 2022 15:39
 Operator : JC\MD
 Sample : VN1114WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1114WBSD01

Quant Time: Nov 15 03:52:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	110139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	194813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	179367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86817	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	65948	44.440	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.880%
35) Dibromofluoromethane	8.171	113	55676	48.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	10.565	98	200565	45.202	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.400%
62) 4-Bromofluorobenzene	12.847	95	74769	47.970	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.940%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.136	85	24000	16.751	ug/l	92
3) Chloromethane	2.377	50	25876	17.003	ug/l	99
4) Vinyl Chloride	2.524	62	34273	15.786	ug/l	94
5) Bromomethane	2.942	94	27501	14.217	ug/l	98
6) Chloroethane	3.124	64	25628	16.462	ug/l	99
7) Trichlorofluoromethane	3.500	101	42439	19.135	ug/l	97
8) Diethyl Ether	3.971	74	16692	20.634	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.383	101	23569	19.243	ug/l	94
10) Methyl Iodide	4.606	142	36314	19.915	ug/l	96
11) Tert butyl alcohol	5.536	59	28552	82.503	ug/l	96
12) 1,1-Dichloroethene	4.353	96	22626	19.380	ug/l	88
13) Acrolein	4.189	56	17153	60.597	ug/l	99
14) Allyl chloride	5.030	41	33822	19.835	ug/l	95
15) Acrylonitrile	5.730	53	71562	97.889	ug/l	99
16) Acetone	4.441	43	63507	101.068	ug/l	96
17) Carbon Disulfide	4.724	76	56229	19.152	ug/l	95
18) Methyl Acetate	5.041	43	41835	20.382	ug/l	99
19) Methyl tert-butyl Ether	5.806	73	85259	19.648	ug/l	100
20) Methylene Chloride	5.283	84	28237	20.155	ug/l	94
21) trans-1,2-Dichloroethene	5.788	96	25727	20.126	ug/l	97
22) Diisopropyl ether	6.683	45	75140	19.968	ug/l	95
23) Vinyl Acetate	6.612	43	257154	81.837	ug/l	98
24) 1,1-Dichloroethane	6.577	63	45175	18.982	ug/l	98
25) 2-Butanone	7.494	43	97113	98.258	ug/l	99
26) 2,2-Dichloropropane	7.500	77	39756	18.158	ug/l	100
27) cis-1,2-Dichloroethene	7.494	96	30216	19.794	ug/l	94
28) Bromochloromethane	7.818	49	18434	18.375	ug/l	100
29) Tetrahydrofuran	7.841	42	58883	92.198	ug/l	99
30) Chloroform	7.971	83	49509	19.160	ug/l	96
31) Cyclohexane	8.259	56	40471	18.337	ug/l	94
32) 1,1,1-Trichloroethane	8.177	97	44746	19.069	ug/l	93
36) 1,1-Dichloropropene	8.377	75	34622	19.554	ug/l	99
37) Ethyl Acetate	7.571	43	37100	19.106	ug/l	98
38) Carbon Tetrachloride	8.365	117	38958	20.003	ug/l	98
39) Methylcyclohexane	9.600	83	46039	19.696	ug/l	94
40) Benzene	8.612	78	111653	20.595	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
 Data File : VN075266.D
 Acq On : 14 Nov 2022 15:39
 Operator : JC\MD
 Sample : VN1114WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1114WBSD01

Quant Time: Nov 15 03:52:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	18490	19.876	ug/l	95
42) 1,2-Dichloroethane	8.677	62	38790	19.667	ug/l	99
43) Isopropyl Acetate	8.694	43	58830	19.532	ug/l	98
44) Trichloroethene	9.353	130	28376	20.863	ug/l	97
45) 1,2-Dichloropropane	9.624	63	27550	21.354	ug/l	97
46) Dibromomethane	9.706	93	20713	20.607	ug/l	97
47) Bromodichloromethane	9.888	83	40033	20.567	ug/l	95
48) Methyl methacrylate	9.682	41	27093	19.610	ug/l	96
49) 1,4-Dioxane	9.694	88	14955	413.361	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	190095	97.806	ug/l	98
52) Toluene	10.629	92	72832	20.098	ug/l	96
53) t-1,3-Dichloropropene	10.835	75	40335	19.632	ug/l	93
54) cis-1,3-Dichloropropene	10.312	75	45103	20.698	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	29658	20.640	ug/l	98
56) Ethyl methacrylate	10.876	69	43180	19.700	ug/l	99
57) 1,3-Dichloropropane	11.165	76	48646	20.329	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	77034	90.911	ug/l	98
59) 2-Hexanone	11.194	43	140297	95.108	ug/l	99
60) Dibromochloromethane	11.359	129	31054	20.484	ug/l	98
61) 1,2-Dibromoethane	11.470	107	30337	20.601	ug/l	99
64) Tetrachloroethene	11.106	164	25964	23.611	ug/l	89
65) Chlorobenzene	11.894	112	76822	20.065	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	28091	19.779	ug/l	99
67) Ethyl Benzene	11.965	91	137880	20.017	ug/l	99
68) m/p-Xylenes	12.070	106	110730	40.063	ug/l	99
69) o-Xylene	12.400	106	53426	19.546	ug/l	98
70) Styrene	12.412	104	86615	19.874	ug/l	99
71) Bromoform	12.576	173	22774	21.417	ug/l #	97
73) Isopropylbenzene	12.694	105	137799	19.771	ug/l	98
74) N-amyl acetate	12.494	43	48197	18.697	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	44982	20.091	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	56594	30.702	ug/l	71
77) Bromobenzene	12.976	156	31458	20.498	ug/l	94
78) n-propylbenzene	13.035	91	158233	20.221	ug/l	99
79) 2-Chlorotoluene	13.123	91	97226	20.333	ug/l	100
80) 1,3,5-Trimethylbenzene	13.176	105	118557	20.252	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.741	75	14259	19.857	ug/l	95
82) 4-Chlorotoluene	13.123	91	97226	20.333	ug/l	100
83) tert-Butylbenzene	13.435	119	104471	20.084	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	122633	20.704	ug/l	100
85) sec-Butylbenzene	13.617	105	149917	20.602	ug/l	99
86) p-Isopropyltoluene	13.729	119	124139	20.657	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	62527	20.809	ug/l	98
88) 1,4-Dichlorobenzene	13.811	146	61839	20.795	ug/l	98
89) n-Butylbenzene	14.053	91	101790	21.033	ug/l	99
90) Hexachloroethane	14.335	117	20278	18.403	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	62550	20.939	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.723	75	9586	20.390	ug/l	95
93) 1,2,4-Trichlorobenzene	15.394	180	29908	23.267	ug/l	99
94) Hexachlorobutadiene	15.500	225	14444	23.049	ug/l	99
95) Naphthalene	15.641	128	107545	21.756	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	28448	22.273	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
Data File : VN075266.D
Acq On : 14 Nov 2022 15:39
Operator : JC\MD
Sample : VN1114WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01

Quant Time: Nov 15 03:52:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 11
Delta R.T. -0.001 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01







Tgt Ion:168 Resp: 110139
Ion Ratio Lower Upper
168 100
99 63.2 51.6 77.4



#2
Dichlorodifluoromethane
Concen: 16.751 ug/l
RT: 2.136 min Scan# 14
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 85 Resp: 24000
Ion Ratio Lower Upper
85 100
87 28.8 16.7 50.0









#3
Chloromethane
Concen: 17.003 ug/l
RT: 2.377 min Scan# 51
Delta R.T. 0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 50 Resp: 25876
Ion Ratio Lower Upper
50 100
52 32.7 25.8 38.6



#4
Vinyl Chloride
Concen: 15.786 ug/l
RT: 2.524 min Scan# 80
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 62 Resp: 34273
Ion Ratio Lower Upper
62 100
64 33.3 24.0 36.0









#5
Bromomethane
Concen: 14.217 ug/l
RT: 2.942 min Scan# 11
Delta R.T. 0.024 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 94 Resp: 27501
Ion Ratio Lower Upper
94 100
96 87.9 71.8 107.6



#6
Chloroethane
Concen: 16.462 ug/l
RT: 3.124 min Scan# 182
Delta R.T. 0.029 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 64 Resp: 25628
Ion Ratio Lower Upper
64 100
66 29.0 23.8 35.8









#7
Trichlorofluoromethane
Concen: 19.135 ug/l
RT: 3.500 min Scan# 244
Delta R.T. 0.011 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01

Tgt Ion:101 Resp: 42439
Ion Ratio Lower Upper
101 100
103 65.1 50.5 75.7




#8
Diethyl Ether
Concen: 20.634 ug/l
RT: 3.971 min Scan# 326
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39


Tgt Ion: 74 Resp: 16692
Ion Ratio Lower Upper
74 100
45 75.2 39.3 117.9




Abundance Scan 396 (4.383 min): VN075019.D\data.ms (-38



m/z--> Scan 396 (4.383 min): VN075266.D\data.ms



Abundance Scan 396 (4.383 min): VN075266.D\data.ms (-34



m/z-->

#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.243 ug/l

RT: 4.383 min Scan# 396

Delta R.T. -0.000 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :

VN1114WBSD01


Tgt Ion:101 Resp: 23569

Ion Ratio Lower Upper


101 100

85 48.7 37.6 56.4


151 79.4 58.3 87.5




Abundance Scan 432 (4.595 min): VN075019.D\data.ms (-42



m/z--> Scan 434 (4.606 min): VN075266.D\data.ms



Abundance Scan 434 (4.606 min): VN075266.D\data.ms (-38



m/z-->

#10

Methyl Iodide

Concen: 19.915 ug/l

RT: 4.606 min Scan# 434

Delta R.T. 0.011 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39


Tgt Ion:142 Resp: 36314


Ion Ratio Lower Upper

142 100

127 45.3 38.0 57.0

141 13.3 12.2 18.4









#11
Tert butyl alcohol
Concen: 82.503 ug/l
RT: 5.536 min Scan# 591
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 59 Resp: 28552
Ion Ratio Lower Upper
59 100
57 9.2 8.5 12.7



#12
1,1-Dichloroethene
Concen: 19.380 ug/l
RT: 4.353 min Scan# 391
Delta R.T. 0.005 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 96 Resp: 22626
Ion Ratio Lower Upper
96 100
61 138.5 126.2 189.4
98 61.6 51.5 77.3









#13
Acrolein
Concen: 60.597 ug/l
RT: 4.189 min Scan# 363
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 56 Resp: 17153
Ion Ratio Lower Upper
56 100
55 68.3 55.2 82.8



#14
Allyl chloride
Concen: 19.835 ug/l
RT: 5.030 min Scan# 506
Delta R.T. 0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 41 Resp: 33822
Ion Ratio Lower Upper
41 100
39 70.6 61.0 91.4
76 39.9 33.4 50.0









#15
Acrylonitrile
Concen: 97.889 ug/l
RT: 5.730 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:	53	Resp:	71562
Ion	Ratio	Lower	Upper
53	100		
52	84.2	66.4	99.6
51	36.2	29.6	44.4



#16
Acetone
Concen: 101.068 ug/l
RT: 4.441 min Scan# 406
Delta R.T. -0.001 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:	43	Resp:	63507
Ion	Ratio	Lower	Upper
43	100		
58	34.4	25.5	38.3









#17
Carbon Disulfide
Concen: 19.152 ug/l
RT: 4.724 min Scan# 453
Delta R.T. 0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 76 Resp: 56229
Ion Ratio Lower Upper
76 100
78 7.5 7.4 11.2



#18
Methyl Acetate
Concen: 20.382 ug/l
RT: 5.041 min Scan# 508
Delta R.T. -0.001 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 43 Resp: 41835
Ion Ratio Lower Upper
43 100
74 28.1 22.0 33.0









#19
Methyl tert-butyl Ether
Concen: 19.648 ug/l
RT: 5.806 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 73 Resp: 85259
Ion Ratio Lower Upper
73 100
57 20.5 16.3 24.5



#20
Methylene Chloride
Concen: 20.155 ug/l
RT: 5.283 min Scan# 549
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 84 Resp: 28237
Ion Ratio Lower Upper
84 100
49 109.9 94.9 142.3
51 35.1 26.8 40.2
86 64.7 53.7 80.5









#21
trans-1,2-Dichloroethene
Concen: 20.126 ug/l
RT: 5.788 min Scan# 611
Delta R.T. -0.012 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:	96	Resp:	25727
Ion	Ratio	Lower	Upper
96	100		
61	122.3	99.9	149.9
98	60.2	51.5	77.3



#22
Diisopropyl ether
Concen: 19.968 ug/l
RT: 6.683 min Scan# 787
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:	45	Resp:	75140
Ion	Ratio	Lower	Upper
45	100		
43	57.3	42.8	64.2
87	33.4	28.4	42.6
59	15.3	10.4	15.6









#23
 Vinyl Acetate
 Concen: 81.837 ug/l
 RT: 6.612 min Scan# 775
 Delta R.T. -0.000 min
 Lab File: VN075266.D
 Acq: 14 Nov 2022 15:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1114WBSD01


Tgt Ion: 43 Resp: 257154
 Ion Ratio Lower Upper
 43 100
 86 12.8 10.7 16.1



#24
 1,1-Dichloroethane
 Concen: 18.982 ug/l
 RT: 6.577 min Scan# 769
 Delta R.T. -0.000 min
 Lab File: VN075266.D
 Acq: 14 Nov 2022 15:39

Tgt Ion: 63 Resp: 45175
 Ion Ratio Lower Upper
 63 100
 98 7.6 3.6 10.8
 100 5.6 2.4 7.2









#25
2-Butanone
Concen: 98.258 ug/l
RT: 7.494 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 43 Resp: 97113
Ion Ratio Lower Upper
43 100
72 32.6 25.5 38.3



#26
2,2-Dichloropropane
Concen: 18.158 ug/l
RT: 7.500 min Scan# 926
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 77 Resp: 39756
Ion Ratio Lower Upper
77 100
97 22.4 11.2 33.6









#27
cis-1,2-Dichloroethene
Concen: 19.794 ug/l
RT: 7.494 min Scan# 91
Delta R.T. 0.005 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 96 Resp: 30216
Ion Ratio Lower Upper
96 100
61 131.1 0.0 278.8
98 66.7 0.0 128.4



#28
Bromochloromethane
Concen: 18.375 ug/l
RT: 7.818 min Scan# 980
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 49 Resp: 18434
Ion Ratio Lower Upper
49 100
129 1.0 0.0 4.6
130 82.6 66.2 99.4









#29
Tetrahydrofuran
Concen: 92.198 ug/l
RT: 7.841 min Scan# 985
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 42 Resp: 58883
Ion Ratio Lower Upper
42 100
72 53.3 43.0 64.6
71 50.9 39.9 59.9



#30
Chloroform
Concen: 19.160 ug/l
RT: 7.971 min Scan# 1006
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 83 Resp: 49509
Ion Ratio Lower Upper
83 100
85 61.2 51.4 77.0









#31
Cyclohexane
Concen: 18.337 ug/l
RT: 8.259 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 56 Resp: 40471
Ion Ratio Lower Upper
56 100
69 37.4 25.2 37.8
84 102.2 77.8 116.8



#32
1,1,1-Trichloroethane
Concen: 19.069 ug/l
RT: 8.177 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 97 Resp: 44746
Ion Ratio Lower Upper
97 100
99 55.3 50.2 75.2
61 49.4 37.0 55.6





#33

1,2-Dichloroethane-d4

Concen: 44.440 ug/l

RT: 8.582 min Scan# 1111

Delta R.T. -0.006 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :





VN1114WBSD01

Tgt Ion: 65 Resp: 65948

Ion Ratio Lower Upper

65 100

67 52.1 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. -0.000 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39




Tgt Ion: 114 Resp: 194813


Ion Ratio Lower Upper

114 100

63 18.2 0.0 39.6

88 14.9 0.0 29.0









#35
Dibromofluoromethane
Concen: 48.683 ug/l
RT: 8.171 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:113 Resp: 55676
Ion Ratio Lower Upper
113 100
111 103.0 80.5 120.7
192 17.7 13.2 19.8



#36
1,1-Dichloropropene
Concen: 19.554 ug/l
RT: 8.377 min Scan# 1075
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 75 Resp: 34622
Ion Ratio Lower Upper
75 100
110 36.3 17.8 53.3
77 30.9 24.6 37.0









#37
Ethyl Acetate
Concen: 19.106 ug/l
RT: 7.571 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:	43	Resp:	37100
Ion	Ratio	Lower	Upper
43	100		
61	13.9	12.2	18.4
70	12.3	9.4	14.2



#38
Carbon Tetrachloride
Concen: 20.003 ug/l
RT: 8.365 min Scan# 1073
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:	117	Resp:	38958
Ion	Ratio	Lower	Upper
117	100		
119	97.4	76.7	115.1
121	32.7	23.9	35.9









#39
Methylcyclohexane
Concen: 19.696 ug/l
RT: 9.600 min Scan# 1115
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 83 Resp: 46039
Ion Ratio Lower Upper
83 100
55 72.2 53.6 80.4
98 47.2 41.1 61.7



#40
Benzene
Concen: 20.595 ug/l
RT: 8.612 min Scan# 1115
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39


Tgt Ion: 78 Resp: 111653
Ion Ratio Lower Upper
78 100
77 23.9 19.4 29.0








#41
Methacrylonitrile
Concen: 19.876 ug/l
RT: 7.782 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01



Tgt Ion: 41 Resp: 18490




Ion	Ratio	Lower	Upper
41	100		
39	59.9	44.4	66.6
67	91.0	69.8	104.6
52	37.5	27.3	40.9




#42
1,2-Dichloroethane
Concen: 19.667 ug/l
RT: 8.677 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 62 Resp: 38790

Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	18.2









#43
Isopropyl Acetate
Concen: 19.532 ug/l
RT: 8.694 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 43 Resp: 58830
Ion Ratio Lower Upper
43 100
61 29.2 22.4 33.6
87 16.3 12.5 18.7



#44
Trichloroethene
Concen: 20.863 ug/l
RT: 9.353 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:130 Resp: 28376
Ion Ratio Lower Upper
130 100
95 95.7 0.0 198.4









#45
1,2-Dichloropropane
Concen: 21.354 ug/l
RT: 9.624 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 63 Resp: 27550
Ion Ratio Lower Upper
63 100
65 31.0 26.1 39.1



#46
Dibromomethane
Concen: 20.607 ug/l
RT: 9.706 min Scan# 1301
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 93 Resp: 20713
Ion Ratio Lower Upper
93 100
95 84.5 66.6 99.8
174 92.7 70.3 105.5









#47
Bromodichloromethane
Concen: 20.567 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 83 Resp: 40033
Ion Ratio Lower Upper
83 100
85 60.8 52.1 78.1
127 7.6 6.6 10.0



#48
Methyl methacrylate
Concen: 19.610 ug/l
RT: 9.682 min Scan# 1297
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 41 Resp: 27093
Ion Ratio Lower Upper
41 100
69 97.2 80.2 120.2
39 65.4 56.2 84.2









#49
1,4-Dioxane
Concen: 413.361 ug/l
RT: 9.694 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 88 Resp: 14955
Ion Ratio Lower Upper
88 100
43 22.8 17.5 26.3
58 59.8 49.8 74.8



#50
Toluene-d8
Concen: 45.202 ug/l
RT: 10.565 min Scan# 1447
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 98 Resp: 200565
Ion Ratio Lower Upper
98 100
100 65.3 53.0 79.6









#51
4-Methyl-2-Pentanone
Concen: 97.806 ug/l
RT: 10.447 min Scan# 1427
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 43 Resp: 190095
Ion Ratio Lower Upper
43 100
58 43.2 33.6 50.4



#52
Toluene
Concen: 20.098 ug/l
RT: 10.629 min Scan# 1458
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 92 Resp: 72832
Ion Ratio Lower Upper
92 100
91 174.2 135.4 203.2









#53
 t-1,3-Dichloropropene
 Concen: 19.632 ug/l
 RT: 10.835 min Scan# 1493
 Delta R.T. -0.006 min
 Lab File: VN075266.D
 Acq: 14 Nov 2022 15:39




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1114WBSD01


Tgt Ion: 75 Resp: 40335
 Ion Ratio Lower Upper
 75 100
 77 28.0 25.4 38.2



#54
 cis-1,3-Dichloropropene
 Concen: 20.698 ug/l
 RT: 10.312 min Scan# 1404
 Delta R.T. -0.006 min
 Lab File: VN075266.D
 Acq: 14 Nov 2022 15:39

Tgt Ion: 75 Resp: 45103
 Ion Ratio Lower Upper
 75 100
 77 29.3 23.7 35.5
 39 43.4 35.8 53.6









#55
1,1,2-Trichloroethane
Concen: 20.640 ug/l
RT: 11.018 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:	97	Resp:	29658
Ion	Ratio	Lower	Upper
97	100		
83	81.3	65.9	98.9
85	47.5	40.4	60.6
99	58.9	46.4	69.6



#56
Ethyl methacrylate
Concen: 19.700 ug/l
RT: 10.876 min Scan# 1500
Delta R.T. -0.001 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:	69	Resp:	43180
Ion	Ratio	Lower	Upper
69	100		
41	56.6	45.6	68.4
39	36.2	28.5	42.7





#57

1,3-Dichloropropane

Concen: 20.329 ug/l

RT: 11.165 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :





VN1114WBSD01

Tgt Ion: 76 Resp: 48646

Ion Ratio Lower Upper

76 100

78 33.2 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 90.911 ug/l

RT: 10.159 min Scan# 1378

Delta R.T. -0.006 min

Lab File: VN075266.D




Acq: 14 Nov 2022 15:39


Tgt Ion: 63 Resp: 77034

Ion Ratio Lower Upper

63 100

106 32.2 24.8 37.2









#59
2-Hexanone
Concen: 95.108 ug/l
RT: 11.194 min Scan# 1554
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 43 Resp: 140297
Ion Ratio Lower Upper
43 100
58 60.1 29.6 88.8



#60
Dibromochloromethane
Concen: 20.484 ug/l
RT: 11.359 min Scan# 1582
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 129 Resp: 31054
Ion Ratio Lower Upper
129 100
127 78.9 38.6 115.8









#61
1,2-Dibromoethane
Concen: 20.601 ug/l
RT: 11.470 min Scan# 11
Delta R.T. -0.001 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:107 Resp: 30337
Ion Ratio Lower Upper
107 100
109 93.5 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 47.970 ug/l
RT: 12.847 min Scan# 1835
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 95 Resp: 74769
Ion Ratio Lower Upper
95 100
174 77.5 0.0 147.4
176 74.1 0.0 142.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:117 Resp: 179367
Ion Ratio Lower Upper
117 100
82 52.8 43.1 64.7
119 33.1 25.5 38.3



#64
Tetrachloroethene
Concen: 23.611 ug/l
RT: 11.106 min Scan# 1539
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:164 Resp: 25964
Ion Ratio Lower Upper
164 100
166 120.0 106.5 159.7
129 93.1 81.9 122.9
131 87.9 81.4 122.2









#65
Chlorobenzene
Concen: 20.065 ug/l
RT: 11.894 min Scan# 1673
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:112 Resp: 76822
Ion Ratio Lower Upper
112 100
114 31.4 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 19.779 ug/l
RT: 11.959 min Scan# 1684
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:131 Resp: 28091
Ion Ratio Lower Upper
131 100
133 97.2 47.8 143.4
119 66.9 33.0 99.0









#67
Ethyl Benzene
Concen: 20.017 ug/l
RT: 11.965 min Scan# 110
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 91 Resp: 137880
Ion Ratio Lower Upper
91 100
106 31.7 24.9 37.3



#68
m/p-Xylenes
Concen: 40.063 ug/l
RT: 12.070 min Scan# 1703
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:106 Resp: 110730
Ion Ratio Lower Upper
106 100
91 196.6 156.0 234.0









#69
o-Xylene
Concen: 19.546 ug/l
RT: 12.400 min Scan# 1759
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:106 Resp: 53426
Ion Ratio Lower Upper
106 100
91 213.1 105.2 315.5



#70
Styrene
Concen: 19.874 ug/l
RT: 12.412 min Scan# 1761
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:104 Resp: 86615
Ion Ratio Lower Upper
104 100
78 48.2 38.1 57.1
103 54.6 44.6 66.8









#71
Bromoform
Concen: 21.417 ug/l
RT: 12.576 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:173 Resp: 22774
Ion Ratio Lower Upper
173 100
175 46.9 24.6 74.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:152 Resp: 86817
Ion Ratio Lower Upper
152 100
115 61.0 30.2 90.6
150 167.1 0.0 348.4









#73
Isopropylbenzene
Concen: 19.771 ug/l
RT: 12.694 min Scan# 1810
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:105 Resp: 137799
Ion Ratio Lower Upper
105 100
120 27.1 13.1 39.1



#74
N-ethyl acetate
Concen: 18.697 ug/l
RT: 12.494 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 43 Resp: 48197
Ion Ratio Lower Upper
43 100
70 50.2 39.8 59.8
55 25.8 19.9 29.9
61 24.9 21.7 32.5









#75
1,1,2,2-Tetrachloroethane
Concen: 20.091 ug/l
RT: 12.935 min Scan# 1850
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 83 Resp: 44982
Ion Ratio Lower Upper
83 100
131 10.1 5.1 15.3
85 63.9 31.6 94.7



#76
1,2,3-Trichloropropane
Concen: 30.702 ug/l
RT: 12.994 min Scan# 1860
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 75 Resp: 56594
Ion Ratio Lower Upper
75 100
77 114.1 75.6 226.8









#77
Bromobenzene
Concen: 20.498 ug/l
RT: 12.976 min Scan# 1857
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:156 Resp: 31458
Ion Ratio Lower Upper
156 100
77 205.3 108.5 325.5
158 98.3 47.7 143.1



#78
n-propylbenzene
Concen: 20.221 ug/l
RT: 13.035 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 91 Resp: 158233
Ion Ratio Lower Upper
91 100
120 24.0 11.9 35.5





#79

2-Chlorotoluene

Concen: 20.333 ug/l

RT: 13.123 min Scan# 1882

Delta R.T. -0.006 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :





VN1114WBSD01

Tgt Ion: 91 Resp: 97226

Ion Ratio Lower Upper

91 100

126 33.5 16.9 50.7



#80

1,3,5-Trimethylbenzene

Concen: 20.252 ug/l

RT: 13.176 min Scan# 1891

Delta R.T. 0.000 min

Lab File: VN075266.D




Acq: 14 Nov 2022 15:39


Tgt Ion: 105 Resp: 118557

Ion Ratio Lower Upper

105 100

120 49.9 24.7 74.1









#81
trans-1,4-Dichloro-2-butene
Concen: 19.857 ug/l
RT: 12.741 min Scan# 1817
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion: 75 Resp: 14259
Ion Ratio Lower Upper
75 100
53 94.0 78.5 117.7
89 43.4 37.7 56.5



#82
4-Chlorotoluene
Concen: 20.333 ug/l
RT: 13.123 min Scan# 1882
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion: 91 Resp: 97226
Ion Ratio Lower Upper
91 100
126 33.5 16.9 50.7









#83
tert-Butylbenzene
Concen: 20.084 ug/l
RT: 13.435 min Scan# 1935
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:119 Resp: 104471
Ion Ratio Lower Upper
119 100
91 64.4 32.8 98.4
134 27.3 12.3 36.9



#84
1,2,4-Trimethylbenzene
Concen: 20.704 ug/l
RT: 13.482 min Scan# 1943
Delta R.T. 0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:105 Resp: 122633
Ion Ratio Lower Upper
105 100
120 45.5 22.7 68.1









#85
sec-Butylbenzene
Concen: 20.602 ug/l
RT: 13.617 min Scan# 1966
Delta R.T. -0.001 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:105 Resp: 149917
Ion Ratio Lower Upper
105 100
134 19.9 10.2 30.6



#86
p-Isopropyltoluene
Concen: 20.657 ug/l
RT: 13.729 min Scan# 1985
Delta R.T. -0.006 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:119 Resp: 124139
Ion Ratio Lower Upper
119 100
134 26.0 13.1 39.3
91 23.9 11.9 35.5





#87

1,3-Dichlorobenzene

Concen: 20.809 ug/l

RT: 13.729 min Scan# 1985

Delta R.T. -0.006 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :

VN1114WBSD01





Tgt Ion:146 Resp: 62527

Ion Ratio Lower Upper

146 100

111 40.9 20.8 62.5

148 64.0 31.1 93.5



#88

1,4-Dichlorobenzene

Concen: 20.795 ug/l

RT: 13.811 min Scan# 1999

Delta R.T. -0.001 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39




Tgt Ion:146 Resp: 61839


Ion Ratio Lower Upper

146 100

111 41.1 20.1 60.3

148 65.9 31.9 95.7





#89

n-Butylbenzene

Concen: 21.033 ug/l

RT: 14.053 min Scan# 2041

Delta R.T. -0.006 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :

VN1114WBSD01





Tgt Ion: 91 Resp: 101790

Ion Ratio Lower Upper

91 100

92 52.4 26.5 79.4

134 27.3 13.0 39.0



#90

Hexachloroethane

Concen: 18.403 ug/l

RT: 14.335 min Scan# 2088

Delta R.T. -0.000 min

Lab File: VN075266.D




Acq: 14 Nov 2022 15:39


Tgt Ion: 117 Resp: 20278

Ion Ratio Lower Upper

117 100

201 68.5 31.5 94.5





#91

1,2-Dichlorobenzene

Concen: 20.939 ug/l

RT: 14.106 min Scan# 2049

Delta R.T. -0.000 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39

Instrument :

MSVOA_N

ClientSampleId :

VN1114WBSD01





Tgt Ion:146 Resp: 62550

Ion Ratio Lower Upper

146 100

111 42.1 21.3 63.9

148 62.9 31.4 94.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 20.390 ug/l

RT: 14.723 min Scan# 2154

Delta R.T. 0.006 min

Lab File: VN075266.D

Acq: 14 Nov 2022 15:39




Tgt Ion: 75 Resp: 9586


Ion Ratio Lower Upper

75 100

155 86.3 42.1 126.3

157 110.8 51.7 155.1









#93
1,2,4-Trichlorobenzene
Concen: 23.267 ug/l
RT: 15.394 min Scan# 211
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39




Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01


Tgt Ion:180 Resp: 29908
Ion Ratio Lower Upper
180 100
182 95.0 47.8 143.4
145 31.7 16.8 50.3



#94
Hexachlorobutadiene
Concen: 23.049 ug/l
RT: 15.500 min Scan# 2286
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:225 Resp: 14444
Ion Ratio Lower Upper
225 100
223 64.0 32.3 96.9
227 62.6 32.0 96.2









#95
Naphthalene
Concen: 21.756 ug/l
RT: 15.641 min Scan# 2310
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Instrument :
MSVOA_N
ClientSampleId :
VN1114WBSD01

Tgt Ion:128 Resp: 107545
Ion Ratio Lower Upper
128 100
127 13.2 10.4 15.6
129 11.3 8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 22.273 ug/l
RT: 15.841 min Scan# 2344
Delta R.T. -0.000 min
Lab File: VN075266.D
Acq: 14 Nov 2022 15:39

Tgt Ion:180 Resp: 28448
Ion Ratio Lower Upper
180 100
182 96.1 47.3 141.8
145 34.6 17.6 52.9

