

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
 Data File : VN075262.D
 Acq On : 14 Nov 2022 13:52
 Operator : JC\MD
 Sample : N5417-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1731

Quant Time: Nov 15 03:51:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	253207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	458698	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	395682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	176743	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	171951	50.401	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.800%
35) Dibromofluoromethane	8.165	113	150829	56.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.020%
50) Toluene-d8	10.565	98	513863	49.186	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.380%
62) 4-Bromofluorobenzene	12.847	95	176404	48.067	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
Data File : VN075262.D
Acq On : 14 Nov 2022 13:52
Operator : JC\MD
Sample : N5417-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1731

Quant Time: Nov 15 03:51:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

