

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
 Data File : VN075263.D
 Acq On : 14 Nov 2022 14:16
 Operator : JC\MD
 Sample : N5417-02
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1732

Quant Time: Nov 15 03:51:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	198562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	356075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	137719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	130473	48.768	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.540%
35) Dibromofluoromethane	8.171	113	115680	55.340	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.680%
50) Toluene-d8	10.565	98	385992	47.595	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.180%
62) 4-Bromofluorobenzene	12.847	95	136109	47.776	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.560%
Target Compounds						
					Qvalue	
18) Methyl Acetate	5.047	43	7074	1.912	ug/l	95
20) Methylene Chloride	5.271	84	4678	1.852	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
Data File : VN075263.D
Acq On : 14 Nov 2022 14:16
Operator : JC\MD
Sample : N5417-02
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1732

Quant Time: Nov 15 03:51:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

