

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
 Data File : VN075264.D
 Acq On : 14 Nov 2022 14:40
 Operator : JC\MD
 Sample : N5572-01 10X
 Misc : 1.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31604

Quant Time: Nov 15 03:52:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	128999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	234535	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	206831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	98236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	87867	50.553	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.100%	
35) Dibromofluoromethane	8.171	113	76619	55.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.300%	
50) Toluene-d8	10.565	98	256910	48.095	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.180%	
62) 4-Bromofluorobenzene	12.847	95	98454	52.468	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.940%	
Target Compounds						Qvalue
39) Methylcyclohexane	9.600	83	3160	1.123	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
Data File : VN075264.D
Acq On : 14 Nov 2022 14:40
Operator : JC\MD
Sample : N5572-01 10X
Misc : 1.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31604

Quant Time: Nov 15 03:52:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

