

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
 Data File : VN075268.D
 Acq On : 14 Nov 2022 16:27
 Operator : JC\MD
 Sample : N5415-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R-28

Quant Time: Nov 15 03:52:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	139818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	252349	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	219253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	95361	50.619	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.240%	
35) Dibromofluoromethane	8.171	113	77845	52.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.100%	
50) Toluene-d8	10.565	98	298206	51.885	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.760%	
62) 4-Bromofluorobenzene	12.847	95	94817	46.963	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.920%	
Target Compounds						
20) Methylene Chloride	5.283	84	16049	9.024	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111422\
Data File : VN075268.D
Acq On : 14 Nov 2022 16:27
Operator : JC\MD
Sample : N5415-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R-28

Quant Time: Nov 15 03:52:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

