

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
 Data File : VN080084.D
 Acq On : 14 Nov 2023 13:16
 Operator : JC\MD
 Sample : PB157075ZHE#18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157075ZHE#18

Quant Time: Nov 15 03:55:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	163155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	308193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	341642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	140892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	154152	62.015	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 124.020%		
35) Dibromofluoromethane	8.171	113	118426	54.292	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.580%		
50) Toluene-d8	10.571	98	437583	54.493	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.980%		
62) 4-Bromofluorobenzene	12.847	95	165846	62.274	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 124.540%#		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	23203	26.574	ug/l	96
20) Methylene Chloride	5.283	84	7159	3.249	ug/l #	88
43) Isopropyl Acetate	8.688	43	126349	25.224	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
Data File : VN080084.D
Acq On : 14 Nov 2023 13:16
Operator : JC\MD
Sample : PB157075ZHE#18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157075ZHE#18

Quant Time: Nov 15 03:55:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

