

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
 Data File : VN080088.D
 Acq On : 14 Nov 2023 14:51
 Operator : JC\MD
 Sample : 05386-01 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CONTAMINATED-WATER

Quant Time: Nov 15 03:56:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	328289	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	346420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	154193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	152484	58.668	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.340%			
35) Dibromofluoromethane	8.171	113	119141	51.277	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.560%			
50) Toluene-d8	10.571	98	417370	48.795	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.580%			
62) 4-Bromofluorobenzene	12.853	95	164102	57.847	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 115.700%			
Target Compounds						
17) Carbon Disulfide	4.706	76	13530	2.554	ug/l #	95
52) Toluene	10.635	92	1420	0.241	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
Data File : VN080088.D
Acq On : 14 Nov 2023 14:51
Operator : JC\MD
Sample : 05386-01 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CONTAMINATED-WATER

Quant Time: Nov 15 03:56:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

