

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
 Data File : VN080089.D
 Acq On : 14 Nov 2023 15:15
 Operator : JC\MD
 Sample : 05386-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CONTAMINATED-PLASTIC

Quant Time: Nov 15 03:56:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	186604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	350177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	376845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	169455	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	155568	54.720	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.440%
35) Dibromofluoromethane	8.171	113	122423	49.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	10.571	98	445661	48.845	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.700%
62) 4-Bromofluorobenzene	12.853	95	179535	59.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.660%

Target Compounds						Qvalue
16) Acetone	4.436	43	46332	46.396	ug/l	97
17) Carbon Disulfide	4.718	76	46601	8.043	ug/l	100
20) Methylene Chloride	5.283	84	7520	2.984	ug/l	91
40) Benzene	8.606	78	71649	6.760	ug/l	99
43) Isopropyl Acetate	8.688	43	132339	23.252	ug/l #	82
52) Toluene	10.635	92	49252	7.838	ug/l	99
67) Ethyl Benzene	11.965	91	11621	0.816	ug/l	99
68) m/p-Xylenes	12.071	106	16443	3.099	ug/l	85
69) o-Xylene	12.400	106	9366	1.833	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	6648	0.675	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
Data File : VN080089.D
Acq On : 14 Nov 2023 15:15
Operator : JC\MD
Sample : 05386-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CONTAMINATED-PLASTIC

Quant Time: Nov 15 03:56:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

