

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
 Data File : VN080093.D
 Acq On : 14 Nov 2023 16:50
 Operator : JC\MD
 Sample : 05376-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 223

Quant Time: Nov 15 03:57:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	180060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	346602	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	384045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	162129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	159174	58.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 116.040%	
35) Dibromofluoromethane	8.171	113	125721	51.250	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.500%	
50) Toluene-d8	10.571	98	458237	50.742	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.480%	
62) 4-Bromofluorobenzene	12.853	95	177608	59.300	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 118.600%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	24790	25.726	ug/l	91
20) Methylene Chloride	5.277	84	9227	3.795	ug/l #	85
43) Isopropyl Acetate	8.694	43	125443m	22.268	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
Data File : VN080093.D
Acq On : 14 Nov 2023 16:50
Operator : JC\MD
Sample : 05376-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
223

Quant Time: Nov 15 03:57:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

