

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
 Data File : VN080100.D
 Acq On : 14 Nov 2023 19:36
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 15 04:45:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	59	0.00
2 T	Dichlorodifluoromethane	0.602	0.578	4.0	55	0.00
3 P	Chloromethane	0.837	0.410	51.0#	32#	0.00
4 C	Vinyl Chloride	0.788	0.725	8.0#	54	0.00
5 T	Bromomethane	0.479	0.312	34.9#	40#	-0.01
6 T	Chloroethane	0.623	0.321	48.5#	32#	0.00
7 T	Trichlorofluoromethane	0.981	1.073	-9.4	65	0.00
8 T	Diethyl Ether	0.372	0.341	8.3	53	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.568	0.513	9.7	55	0.00
10 T	Methyl Iodide	0.649	0.565	12.9	48#	0.00
11 T	Tert butyl alcohol	0.097	0.093	4.1	59	0.00
12 CM	1,1-Dichloroethene	0.538	0.466	13.4#	50	-0.01
13 T	Acrolein	0.080	0.126	-57.5#	95	-0.01
14 T	Allyl chloride	0.824	0.708	14.1	51	0.00
15 T	Acrylonitrile	0.303	0.258	14.9	51	0.00
16 T	Acetone	0.268	0.225	16.0	53	0.00
17 T	Carbon Disulfide	1.552	1.050	32.3#	41#	0.00
18 T	Methyl Acetate	1.280	1.031	19.5	50	0.00
19 T	Methyl tert-butyl Ether	1.853	1.942	-4.8	60	0.00
20 T	Methylene Chloride	0.675	0.556	17.6	51	0.01
21 T	trans-1,2-Dichloroethene	0.600	0.506	15.7	51	0.00
22 T	Diisopropyl ether	1.910	1.615	15.4	48#	0.00
23 T	Vinyl Acetate	1.054	0.837	20.6	44#	0.00
24 P	1,1-Dichloroethane	1.161	1.054	9.2	54	0.00
25 T	2-Butanone	0.392	0.312	20.4	46#	0.00
26 T	2,2-Dichloropropane	0.960	1.005	-4.7	60	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.637	11.3	53	0.00
28 T	Bromochloromethane	0.566	0.516	8.8	54	0.00
29 T	Tetrahydrofuran	0.255	0.196	23.1	44#	0.00
30 C	Chloroform	1.261	1.213	3.8#	60	0.00
31 T	Cyclohexane	1.007	0.747	25.8#	46#	0.00
32 T	1,1,1-Trichloroethane	1.028	1.160	-12.8	67	0.00
33 S	1,2-Dichloroethane-d4	0.762	0.847	-11.2	64	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	55	0.00
35 S	Dibromofluoromethane	0.354	0.375	-5.9	57	0.00
36 T	1,1-Dichloropropene	0.505	0.495	2.0	54	0.00
37 T	Ethyl Acetate	0.463	0.399	13.8	46#	0.00
38 T	Carbon Tetrachloride	0.509	0.633	-24.4	68	0.00
39 T	Methylcyclohexane	0.476	0.449	5.7	49#	0.00
40 TM	Benzene	1.513	1.430	5.5	52	0.00
41 T	Methacrylonitrile	0.246	0.221	10.2	48#	0.00
42 TM	1,2-Dichloroethane	0.529	0.627	-18.5	66	0.00
43 T	Isopropyl Acetate	0.813	0.780	4.1	53	0.00
44 TM	Trichloroethene	0.376	0.352	6.4	52	0.00
45 C	1,2-Dichloropropane	0.387	0.362	6.5#	51	0.00
46 T	Dibromomethane	0.264	0.264	0.0	55	0.00
47 T	Bromodichloromethane	0.563	0.607	-7.8	61	0.00
48 T	Methyl methacrylate	0.356	0.341	4.2	48#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
 Data File : VN080100.D
 Acq On : 14 Nov 2023 19:36
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 15 04:45:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	55	0.00
50 S	Toluene-d8	1.303	1.316	-1.0	53	0.00
51 T	4-Methyl-2-Pentanone	0.458	0.410	10.5	47#	0.00
52 CM	Toluene	0.897	0.918	-2.3#	53	0.00
53 T	t-1,3-Dichloropropene	0.552	0.631	-14.3	59	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.645	-7.1	54	0.00
55 T	1,1,2-Trichloroethane	0.361	0.365	-1.1	55	0.00
56 T	Ethyl methacrylate	0.515	0.561	-8.9	54	0.00
57 T	1,3-Dichloropropane	0.636	0.654	-2.8	56	0.00
58 T	2-Chloroethyl Vinyl ether	0.261	0.263	-0.8	50	0.00
59 T	2-Hexanone	0.350	0.303	13.4	45#	0.00
60 T	Dibromochloromethane	0.386	0.455	-17.9	62	0.00
61 T	1,2-Dibromoethane	0.362	0.361	0.3	53	0.00
62 S	4-Bromofluorobenzene	0.432	0.519	-20.1	61	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	54	0.00
64 T	Tetrachloroethene	0.324	0.333	-2.8	55	0.00
65 PM	Chlorobenzene	1.085	1.077	0.7	54	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.453	-15.9	61	0.00
67 C	Ethyl Benzene	1.890	2.004	-6.0#	54	0.00
68 T	m/p-Xylenes	0.704	0.746	-6.0	54	0.00
69 T	o-Xylene	0.678	0.732	-8.0	55	0.00
70 T	Styrene	1.089	1.265	-16.2	56	0.00
71 P	Bromoform	0.277	0.345	-24.5	61	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
73 T	Isopropylbenzene	3.795	3.842	-1.2	58	0.00
74 T	N-amyl acetate	1.656	1.273	23.1	43#	0.00
75 P	1,1,2,2-Tetrachloroethane	1.345	1.169	13.1	53	0.00
76 T	1,2,3-Trichloropropane	1.174	0.949	19.2	50#	0.00
77 T	Bromobenzene	0.929	0.909	2.2	58	0.00
78 T	n-propylbenzene	4.346	4.416	-1.6	56	0.00
79 T	2-Chlorotoluene	2.776	2.851	-2.7	60	0.00
80 T	1,3,5-Trimethylbenzene	2.952	3.315	-12.3	62	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.361	16.2	45#	0.00
82 T	4-Chlorotoluene	2.759	2.841	-3.0	59	0.00
83 T	tert-Butylbenzene	2.417	2.705	-11.9	61	0.00
84 T	1,2,4-Trimethylbenzene	2.905	3.309	-13.9	61	0.00
85 T	sec-Butylbenzene	3.348	3.647	-8.9	59	0.00
86 T	p-Isopropyltoluene	2.730	3.121	-14.3	61	0.00
87 T	1,3-Dichlorobenzene	1.728	1.689	2.3	58	0.00
88 T	1,4-Dichlorobenzene	1.754	1.663	5.2	57	0.00
89 T	n-Butylbenzene	2.351	2.550	-8.5	58	0.00
90 T	Hexachloroethane	0.541	0.588	-8.7	67	0.00
91 T	1,2-Dichlorobenzene	1.632	1.644	-0.7	59	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.228	0.265	-16.2	65	0.00
93 T	1,2,4-Trichlorobenzene	0.776	0.888	-14.4	63	0.00
94 T	Hexachlorobutadiene	0.352	0.460	-30.7#	79	0.00
95 T	Naphthalene	2.631	2.625	0.2	55	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111423\
Data File : VN080100.D
Acq On : 14 Nov 2023 19:36
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 15 04:45:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.760	0.838	-10.3	61	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6