

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084843.D
 Acq On : 14 Nov 2024 10:28
 Operator : JC\MD
 Sample : VN1114MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1114MBL01

Quant Time: Nov 15 00:42:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	167986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	292808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105784	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	120650	49.726	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.460%
35) Dibromofluoromethane	8.159	113	96565	48.722	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	10.565	98	334898	45.871	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.740%
62) 4-Bromofluorobenzene	12.847	95	120812	44.276	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084843.D
Acq On : 14 Nov 2024 10:28
Operator : JC\MD
Sample : VN1114MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1114MBL01

Quant Time: Nov 15 00:42:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

