

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084849.D
 Acq On : 14 Nov 2024 14:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Nov 15 00:47:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	211189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	364075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	313468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	144718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	145589	47.730	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.460%		
35) Dibromofluoromethane	8.159	113	117217	47.565	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.140%		
50) Toluene-d8	10.565	98	421705	46.454	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.900%		
62) 4-Bromofluorobenzene	12.847	95	157235	46.345	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.700%		
Target Compounds						
95) Naphthalene	15.635	128	9293	1.109	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084849.D
Acq On : 14 Nov 2024 14:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 15 00:47:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

