

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084850.D
 Acq On : 14 Nov 2024 14:36
 Operator : JC\MD
 Sample : P4844-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 241113050-06-TRIP-BLANK

Quant Time: Nov 15 00:47:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	192705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	332008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	140233	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134380	48.281	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	96.560%	
35) Dibromofluoromethane	8.159	113	110410	49.130	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.260%	
50) Toluene-d8	10.565	98	387856	46.852	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	93.700%	
62) 4-Bromofluorobenzene	12.847	95	149514	48.326	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	96.660%	
Target Compounds						
95) Naphthalene	15.635	128	3268	0.403	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084850.D
Acq On : 14 Nov 2024 14:36
Operator : JC\MD
Sample : P4844-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
241113050-06-TRIP-BLANK

Quant Time: Nov 15 00:47:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

