

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084852.D
 Acq On : 14 Nov 2024 15:23
 Operator : JC\MD
 Sample : P4819-01 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1594

Quant Time: Nov 15 00:48:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	199541	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	341153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	148640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	136920	47.508	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	95.020%	
35) Dibromofluoromethane	8.159	113	109594	47.460	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.920%	
50) Toluene-d8	10.565	98	400791	47.117	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.240%	
62) 4-Bromofluorobenzene	12.847	95	157057	49.403	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.800%	
Target Compounds						
16) Acetone	4.424	43	93151	110.297	ug/l	98
25) 2-Butanone	7.482	43	26890	19.999	ug/l #	87
51) 4-Methyl-2-Pentanone	10.441	43	18610	6.718	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084852.D
Acq On : 14 Nov 2024 15:23
Operator : JC\MD
Sample : P4819-01 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1594

Quant Time: Nov 15 00:48:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

