

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084853.D
 Acq On : 14 Nov 2024 15:47
 Operator : JC\MD
 Sample : P4718-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WB-307-SW

Quant Time: Nov 15 00:48:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	203676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	345476	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	299825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	136300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	140566	47.783	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.560%
35) Dibromofluoromethane	8.159	113	112509	48.113	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.220%
50) Toluene-d8	10.565	98	397675	46.166	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.340%
62) 4-Bromofluorobenzene	12.847	95	148708	46.192	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.380%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084853.D
Acq On : 14 Nov 2024 15:47
Operator : JC\MD
Sample : P4718-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WB-307-SW

Quant Time: Nov 15 00:48:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

