

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084854.D
 Acq On : 14 Nov 2024 16:11
 Operator : JC\MD
 Sample : PB164945TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164945TB

Quant Time: Nov 15 00:49:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	180574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	307757	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121421	46.555	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.120%		
35) Dibromofluoromethane	8.159	113	98502	47.285	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.580%		
50) Toluene-d8	10.565	98	368009	47.958	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.920%		
62) 4-Bromofluorobenzene	12.847	95	142593	49.721	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.440%		
Target Compounds						
16) Acetone	4.430	43	24938	30.598	ug/l	94
20) Methylene Chloride	5.265	84	10821	4.717	ug/l	86
43) Isopropyl Acetate	8.688	43	92362	18.464	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084854.D
Acq On : 14 Nov 2024 16:11
Operator : JC\MD
Sample : PB164945TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164945TB

Quant Time: Nov 15 00:49:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

