

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084856.D
 Acq On : 14 Nov 2024 16:59
 Operator : JC\MD
 Sample : P4821-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-F-24

Quant Time: Nov 15 00:50:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	177056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	308645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126089	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	121957	47.690	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.380%		
35) Dibromofluoromethane	8.165	113	97875	46.849	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.700%		
50) Toluene-d8	10.565	98	371601	48.286	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.580%		
62) 4-Bromofluorobenzene	12.847	95	138579	48.182	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.360%		
Target Compounds						
16) Acetone	4.424	43	47183	61.725	ug/l	97
20) Methylene Chloride	5.271	84	4160	1.850	ug/l #	75
43) Isopropyl Acetate	8.688	43	120687	24.356	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084856.D
Acq On : 14 Nov 2024 16:59
Operator : JC\MD
Sample : P4821-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-F-24

Quant Time: Nov 15 00:50:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

