

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084857.D
 Acq On : 14 Nov 2024 17:23
 Operator : JC\MD
 Sample : P4821-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-F-2

Quant Time: Nov 15 00:50:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	175039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	126515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	122716	48.540	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.080%
35) Dibromofluoromethane	8.165	113	98473	47.252	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.500%
50) Toluene-d8	10.565	98	363263	47.320	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.640%
62) 4-Bromofluorobenzene	12.847	95	143644	50.066	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.140%
Target Compounds						
16) Acetone	4.424	43	50087	66.491	ug/l	97
20) Methylene Chloride	5.265	84	4499	2.023	ug/l #	86
43) Isopropyl Acetate	8.688	43	125878	25.511	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084857.D
Acq On : 14 Nov 2024 17:23
Operator : JC\MD
Sample : P4821-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-F-2

Quant Time: Nov 15 00:50:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

