

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
 Data File : VN084863.D
 Acq On : 14 Nov 2024 19:46
 Operator : JC\MD
 Sample : P4770-03 500X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-10

Quant Time: Nov 15 00:54:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	170053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	300668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118659	48.311	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.620%		
35) Dibromofluoromethane	8.165	113	95429	46.890	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.780%		
50) Toluene-d8	10.565	98	346208	46.180	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.360%		
62) 4-Bromofluorobenzene	12.847	95	135390	48.322	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.640%		
Target Compounds						
52) Toluene	10.629	92	1518	0.286	ug/l	80
68) m/p-Xylenes	12.065	106	1983	0.536	ug/l #	71
78) n-propylbenzene	13.029	91	3937	0.389	ug/l	97
80) 1,3,5-Trimethylbenzene	13.170	105	2778	0.388	ug/l	93
84) 1,2,4-Trimethylbenzene	13.482	105	9830	1.330	ug/l	96
95) Naphthalene	15.635	128	3240	0.454	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111424\
Data File : VN084863.D
Acq On : 14 Nov 2024 19:46
Operator : JC\MD
Sample : P4770-03 500X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-10

Quant Time: Nov 15 00:54:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

