

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052333.D
 Acq On : 15 Nov 2018 10:50
 Operator : MD\SY
 Sample : PB114830TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114830TB

Quant Time: Nov 16 03:18:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	344907	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	576654	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	494098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	235357	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
35) Dibromofluoromethane	7.59	113	184440	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
50) Toluene-d8	10.09	98	729120	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
62) 4-Bromofluorobenzene	12.40	95	213637	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	
Target Compounds						
16) Acetone	3.82	43	6872	5.12	ug/l	92
20) Methylene Chloride	4.55	84	13161	2.56	ug/l	92
43) Isopropyl Acetate	8.17	43	20924	2.59	ug/l #	88
80) 1,3,5-Trimethylbenzene	12.73	105	11185	1.04	ug/l	95
95) Naphthalene	15.13	128	2869	3.06	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052333.D
Acq On : 15 Nov 2018 10:50
Operator : MD\SY
Sample : PB114830TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114830TB

Quant Time: Nov 16 03:18:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

