

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052335.D
 Acq On : 15 Nov 2018 11:43
 Operator : MD\SY
 Sample : J5939-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OILY-WATER

Quant Time: Nov 16 03:24:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	322159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	545984	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	442341	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147978	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	222533	48.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.82%	
35) Dibromofluoromethane	7.59	113	171491	48.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.40%	
50) Toluene-d8	10.09	98	667805	48.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.42%	
62) 4-Bromofluorobenzene	12.40	95	188500	41.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.72%	

Target Compounds						Qvalue
16) Acetone	3.82	43	3821	3.05	ug/l	99
73) Isopropylbenzene	12.25	105	14843	1.25	ug/l	94
85) sec-Butylbenzene	13.17	105	44805	3.94	ug/l	94
89) n-Butylbenzene	13.61	91	8430	1.02	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052335.D
Acq On : 15 Nov 2018 11:43
Operator : MD\SY
Sample : J5939-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OILY-WATER

Quant Time: Nov 16 03:24:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

