

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052337.D
 Acq On : 15 Nov 2018 12:36
 Operator : MD\SY
 Sample : PB114828ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#01

Quant Time: Nov 16 03:33:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	325976	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	547140	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	456245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145451	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	226177	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
35) Dibromofluoromethane	7.59	113	176487	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
50) Toluene-d8	10.09	98	687688	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.40	95	189815	41.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.12%	
Target Compounds						
16) Acetone	3.82	43	7050	5.56	ug/l	99
20) Methylene Chloride	4.56	84	14635	3.14	ug/l	94
43) Isopropyl Acetate	8.17	43	23769	3.09	ug/l #	89
80) 1,3,5-Trimethylbenzene	12.73	105	9366	1.01	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052337.D
Acq On : 15 Nov 2018 12:36
Operator : MD\SY
Sample : PB114828ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#01

Quant Time: Nov 16 03:33:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

