

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052338.D
 Acq On : 15 Nov 2018 13:03
 Operator : MD\SY
 Sample : PB114828ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#02

Quant Time: Nov 16 03:35:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	320691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	538239	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	462725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	159519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	223477	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
35) Dibromofluoromethane	7.59	113	170870	49.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.44%	
50) Toluene-d8	10.09	98	681688	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
62) 4-Bromofluorobenzene	12.40	95	195480	43.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.02%	
Target Compounds						
16) Acetone	3.82	43	6925	5.55	ug/l	96
20) Methylene Chloride	4.56	84	13282	2.84	ug/l	95
43) Isopropyl Acetate	8.17	43	24715	3.27	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052338.D
 Acq On : 15 Nov 2018 13:03
 Operator : MD\SY
 Sample : PB114828ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleID :
 PB114828ZHE#02

Quant Time: Nov 16 03:35:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

