

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052339.D
 Acq On : 15 Nov 2018 13:30
 Operator : MD\SY
 Sample : PB114828ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#03

Quant Time: Nov 16 03:44:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	325759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	544386	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	467100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	162594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	225842	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
35) Dibromofluoromethane	7.59	113	177329	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
50) Toluene-d8	10.09	98	687592	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	12.40	95	201485	44.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.68%	
Target Compounds						
16) Acetone	3.82	43	6510	5.14	ug/l	# 87
20) Methylene Chloride	4.55	84	14116	3.00	ug/l	96
43) Isopropyl Acetate	8.17	43	21814	2.85	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052339.D
Acq On : 15 Nov 2018 13:30
Operator : MD\SY
Sample : PB114828ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#03

Quant Time: Nov 16 03:44:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

