

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
 Data File : VN052340.D
 Acq On : 15 Nov 2018 14:26
 Operator : MD\SY
 Sample : PB114828ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114828ZHE#04

Quant Time: Nov 16 03:46:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	316745	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	534483	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	458335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	154281	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	222200	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
35) Dibromofluoromethane	7.59	113	172417	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
50) Toluene-d8	10.09	98	679472	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
62) 4-Bromofluorobenzene	12.40	95	192188	43.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.16%	
Target Compounds						
16) Acetone	3.81	43	6888	5.59	ug/l #	84
20) Methylene Chloride	4.56	84	13472	2.94	ug/l	98
43) Isopropyl Acetate	8.17	43	25374	3.38	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111518\
Data File : VN052340.D
Acq On : 15 Nov 2018 14:26
Operator : MD\SY
Sample : PB114828ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114828ZHE#04

Quant Time: Nov 16 03:46:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

